



Application Note AN-NIR-129

Dried pet food analysis by near-infrared spectroscopy (NIRS)

Rapid determination of moisture, protein, fat, and ash

Pet food testing involves accurate monitoring and control of moisture, protein, fat, and ash content. This allows pet food manufacturers to satisfy nutritional adequacy requirements, meet regulatory guidelines, and provide high-quality foods to pet owners.

Traditional analysis methods are often slow and labor-intensive. Near-infrared (NIR) spectroscopy provides a

fast, nondestructive alternative for analyzing these key quality control parameters with minimal sample preparation. This study evaluates the use of NIR spectroscopy in the pet food industry for efficient and reliable quality control during dried pet food production.

EXPERIMENTAL EQUIPMENT

199 dried pet food samples (mainly dog and cat food) from different suppliers were analyzed in whole and ground forms on an OMNIS NIR Analyzer Solid (Figure 1). The samples were added to an OMNIS sample cup and analyzed in diffuse reflection mode. To include sample variety, the sample rotated during measurement to collect spectra from different locations. The automatically averaged spectra were used for model development. Reference values were obtained by official methods such as ISO 6496 (moisture), ISO 5983 (protein), ISO 5984 (ash), and ISO 6492 (fat).



Figure 1. OMNIS NIR Analyzer Solid

RESULT

The obtained NIR spectra (Figure 2) were used to create prediction models for the different reference parameters. An external validation set was used to verify the predictive performance of the calculated

prediction models. Correlation diagrams, which display the relation between the NIR prediction and the reference values, are shown in Figures 3–10 together with the respective figures of merit (FOM).




Figure 2. NIR spectra of pet food samples. Data was obtained with an OMNIS NIR Analyzer Solid. Spectra shown in blue have been used to calibrate the model, while green spectra have been used for validation.

Result moisture in dried pet food (whole)




Figure 3. Correlation diagram and the respective FOMs for the prediction of moisture in dried pet food (whole) using an OMNIS NIR Analyzer Solid. Blue dots represent calibration samples, green dots represent samples used to validate the model.

Result moisture in dried pet food (ground)




Figure 4. Correlation diagram and the respective FOMs for the prediction of moisture in dried pet food (ground) using an OMNIS NIR Analyzer Solid. Blue dots represent calibration samples, green dots represent samples used to validate the model.

Result protein in dried pet food (whole)




Figure 5. Correlation diagram and the respective FOMs for the prediction of protein in dried pet food (whole) using an OMNIS NIR Analyzer Solid. Blue dots represent calibration samples, green dots represent samples used to validate the model.

Result protein in dried pet food (ground)




Figure 6. Correlation diagram and the respective FOMs for the prediction of protein in dried pet food (ground) using an OMNIS NIR Analyzer Solid. Blue dots represent calibration samples, green dots represent samples used to validate the model.

Result fat in dried pet food (whole)




Figure 7. Correlation diagram and the respective FOMs for the prediction of fat in dried pet food (whole) using an OMNIS NIR Analyzer Solid. Blue dots represent calibration samples, green dots represent samples used to validate the model.

Result fat in dried pet food (ground)




Figure 8. Correlation diagram and the respective FOMs for the prediction of fat in dried pet food (ground) using an OMNIS NIR Analyzer Solid. Blue dots represent calibration samples, green dots represent samples used to validate the model.

RESULT

Result ash in dried pet food (whole)




Figure 9. Correlation diagram and the respective FOMs for the prediction of ash in dried pet food (whole) using an OMNIS NIR Analyzer Solid. Blue dots represent calibration samples, green dots represent samples used to validate the model.

RESULT

Result ash in dried pet food (ground)




Figure 10. Correlation diagram and the respective FOMs for the prediction of ash in dried pet food (ground) using an OMNIS NIR Analyzer Solid. Blue dots represent calibration samples, green dots represent samples used to validate the model.

CONCLUSION

This Application Note demonstrates the feasibility of using NIR spectroscopy for the analysis of several pet food quality parameters, including protein, moisture, fat, and ash. Results indicate that lower prediction errors can be obtained when grinding pet food samples before their analysis. A comparison between different methods used for the analysis of pet food

clearly highlights the significant time-saving potential of NIR spectroscopy (**Table 1**).

Furthermore, the simplicity of the analysis and the availability of a pre-calibrated instrument, using the **pre-calibration, dry pet food, solid (6.06008.027)**, make both the implementation and the application of NIR spectroscopy with OMNIS NIRS straightforward.

Table 1. Overview of time effort for the analysis of different parameters in pet food by traditional methods and NIR spectroscopy.

Parameter	Standard	Process	Estimated time
Moisture	ISO 6496	Oven drying	2.5 hours
Protein	ISO 5983	Kjeldahl method	4.0 hours
Fat	ISO 6492	Soxhlet extraction	4.0 hours
Ash	ISO 5984	Incineration	4.0 hours
Moisture, protein, fat, ash	ASTM D6122	NIR spectroscopy	1.0 minute

CONTACT

Metrohm Inula
Shuttleworthstraße 25
1210 Wien

office@metrohm.at

CONFIGURATION



OMNIS NIR Analyzer Solid

Nahinfrarot-Spektrometer für feste und viskose Proben.

Der OMNIS NIR Analyzer ist die nach Schweizer Qualitätsstandards entwickelte und produzierte Nahinfrarot-Spektroskopie (NIRS) Lösung für die Routineanalytik entlang der gesamten Produktionskette. Die Nutzung neuester Technologien und die Einbindung in die moderne OMNIS Software spiegeln sich in der Geschwindigkeit, der Bedienbarkeit und dem flexiblen Einsatz dieser NIR-Spektrometer wider.

Die Vorteile des OMNIS NIR Analyzer Solid im Überblick:

- Messungen von Feststoffen und viskosen Proben in weniger als 10 Sekunden
- Automatisierte Mehrpositionenmessungen für reproduzierbare Ergebnisse auch bei inhomogenen Proben
- Einfache Einbindung in ein Automationssystem oder Verknüpfung mit weiteren Analysetechnologien (Titration)
- Unterstützung zahlreicher Probengefässe.



Grosser Halter OMNIS NIR, 100 mm

Grosser Halter für Grosses Probengefäss OMNIS NIR, 100 mm (6.07402.110).

Erlaubt eindeutige Positionierung des Probengefässes und die Rotation des Probengefässes.



Grosse Schale OMNIS NIR, 100 mm

Grosses Probengefäss für die Spektrenaufnahme von Pulvern und Granulaten in Reflexion an unterschiedlichen Probenstellen.

Kompatibel mit:

- Grossen Halter OMNIS NIR, 100 mm (6.07402.100)

OMNIS

A WHOLE NEW LEVEL OF PERFORMANCE

OMNIS Stand-Alone Lizenz

Ermöglicht den Stand-Alone Betrieb der OMNIS Software auf einem Windows™ Computer.

Merkmale:

- Die Lizenz enthält bereits eine OMNIS Geräte Lizenz.
- Muss über das Metrohm Lizenzierungsportal aktiviert werden.

Softwarelizenz Quant Development

Softwarelizenz für die Erstellung und Bearbeitung von Quantifizierungsmodellen in einer Stand-Alone OMNIS Software Installation.

OMNIS

A WHOLE NEW LEVEL OF PERFORMANCE



Vorkalibrierung, Trockenfutter für Haustiere, fest
OMNIS Vorkalibrierung für die Bestimmung der Feuchtigkeit, des Protein, Fett und Aschegehalts in Trockenfutter für Haustiere mittels NIR-Spektroskopie.