



Application Note AN-RS-014

Trace Detection of Rhodamine B in Cayenne Powder

Protecting consumer safety with Misa

The addition of dyes to provide uniform coloration and enhance visual appeal in food products is a common practice. Rhodamine B is a dye utilized extensively in biotechnology and industrial applications and is one of several colorants banned for use as food additives in Europe and North America. The most common analytical methods for detection of illicit dyes in food products, GC/MS and HPLC, are laboratory-based instrumental methods that require

specialized training. With Misa (Metrohm Instant SERS Analyzer), detection of trace amounts of Rhodamine B in ground cayenne pepper is quick and easy after a facile extraction procedure with minimal material consumption. Rhodamine B can be detected in cayenne powder at a concentration of 50 $\mu\text{g/g}$. However, a simple concentration step improves that limit to 10 $\mu\text{g/g}$.

INTRODUCTION

Ground cayenne pepper bought commercially was doped with Rhodamine B (RhB) and tested

with Misa to simulate a realistic food screening scenario.

REFERENCE MATERIAL AND LIBRARY CREATION

To establish a reference spectrum, a pure RhB standard (50 $\mu\text{g/g}$ in ultrapure water) was analyzed using gold nanoparticles (Au NPs). The unique SERS spectrum shown in **Figure 1** can be used to create a library entry for RhB.




Figure 1. Gold NP SERS standard reference spectrum of Rhodamine B.

EXPERIMENT

A stock solution of RhB in methanol was prepared. Purchased cayenne powder was treated with serial dilutions of the stock (3 mL stock to 1 g cayenne) to yield samples with 1000, 500, 100, 50, 10, 5, and 1 $\mu\text{g/g}$ of RhB. Samples were thoroughly mixed and air-dried. To prepare extracts, 0.1 g of each spiked sample was added to a vial with 400 μL of methanol, shaken to mix, and left to settle for 10 minutes. To prepare test samples, 50 μL of the methanol extract was pipetted into a vial with 400 μL of Au NP solution and 50 μL of 0.5 mol/L salt solution. The vial was shaken to mix, and then placed into the vial attachment on Misa for testing.



Table 1. Experimental Parameters

Instrument		Acquisition	
Firmware	0.9.33	Laser Power	5
Software	Misa Cal V1.0.15	Int. Time	10 s
Misa Vial Attachment	6.07505.040	Averages	10
ID Kit - Au NP	6.07506.440	Raster	ON

RESULTS

In **Figure 2**, overlaid spectra of RhB indicate detection down to 50 $\mu\text{g/g}$. For each concentration tested, the baseline spectrum

from unadulterated cayenne was subtracted from the average of baseline-corrected, replicate measurements.




Figure 2. Gold NP SERS concentration profile of RhB extracted from adulterated cayenne powder. Spectra are baselined, with Au NP and control subtracted.

To improve trace detection and spectral signal-to-noise, a very simple concentration method was applied to each extract. All extracts were fully air-dried, then resuspended in methanol to

yield a 5x increase in concentration. The spectra in **Figure 3** demonstrate detection of RhB down to 10 $\mu\text{g/g}$.




Figure 3. RhB profile after 5x concentration.

FIELD TEST PROTOCOL

Detection of Rhodamine B in the field

Using the large end of the scoop, add 3–4 scoops of sample to a 2 mL vial. Add methanol to the vial until ~1/3 full. Cap and shake the vial gently to mix, then let sample rest for 2 minutes. Fill a clean vial halfway full with Au NPs. Using

pipettes, add 1 drop each of sample solution and NaCl solution to the Au NPs, then cap and shake the vial gently to mix. Insert into vial attachment on Misa for measurement.

Table 2. Requirements for Field Test Protocol

ID Kit - Au NP	6.07506.440
includes:	Gold nanoparticles (Au NP)
	Scoop
	Disposable pipettes
	2 mL glass vials
Reagents	
Methanol	
NaCl solution	3 g NaCl in 100 mL water
Test settings	Use ID Kit OP on MISA

CONCLUSION

Trace levels of detection, ease of sample preparation, and rapid assay times collectively recommend Misa as a reliable, cost-effective

solution for high-throughput, on-site identification of adulterated food products.

CONTACT

メトロームジャパン株式会社
143-0006 東京都大田区平
和島6-1-1
null 東京流通センター アネ
ックス9階

metrohm.jp@metrohm.jp

CONFIGURATION



MISA Advanced

Metrohm Instant SERS Analyzer (MISA) は、微量レベルでの違法物質、食品添加物、および食品汚染物質の迅速な検出 / 同定のための高性能な携帯可能分析システムです。MISAは、Metrohm 独自の軌道ラスタースキャン技術 (Orbital Raster Scan Technologie, ORS) を備えた高効率の分光器を有しています。これは省スペースで、より長いバッテリー寿命を持ち、現場やラボでの移動式用途にも完璧に適しています。MISA ではフレキシブルなサンプル採取を可能にする、レーザークラス1の様々なアタッチメントをご利用いただけます。アナライザーはBluetoothまたはUSBコネクタを介して操作可能です。

MISA Advanced パッケージは、ユーザーに Metrohmのナノ粒子溶液とP-SERSストリップを用いたSERS分析を可能にするコンプリートパッケージです。

MISA Advanced パッケージには、MISAバイアルアタッチメント、P-SERSアタッチメント、ASTM校正標準、USBミニケーブル、USB電源装置、ならびにMISA装置を操作するためのMISA Calソフトウェアが含まれます。装置と付属品を安全に保管するための頑丈な保護ケースも同梱されています。



ID - Au NP

IDキット - Au NPには、Mira/Misaユーザーが金コロイド溶液でSERS分析を行うのに必要なコンポーネントが含まれています。このキットには、使い捨てのへら、滴下ピペット、サンプルボトル、および金コロイド入りのボトルが含まれています