



Separate Ag ring electrode

6.00350.100

This silver electrode must be used in combination with a reference electrode ($c(\text{KNO}_3) = 1 \text{ mol/L}$ as reference electrolyte) and is suitable for precipitation titrations (titrant silver nitrate), e.g. of:

- chloride, bromide, iodide
- sulfides
- hydrogen sulfide
- mercaptans
- cyanides

The permanently fused-in silver ring is resistant to highly concentrated acids and saline solutions.

Below, the accessories are grouped into Scope of delivery and Optional accessories.

Please keep this printout at hand for ordering replacement material.

These lists may be subject to change.

Scope of delivery 6.00350.100

Qt.	Order no.	Description
-----	-----------	-------------

100 Stk

6.1236.050

Sleeve with SGJ 14/12 mm

Sleeve with SGJ 14/12 mm, polyethylene.



100 Stk

6.2008.040

Storage vessel

Together with 6.2008.050 storage vessel holder. it provides a support for the electrode on 807 Dosing Units.



Optional accessories

Order no.	Description	
6.2104.020	Electrode cable / 1 m / F For connecting electrodes with Metrohm plug-in head G to Metrohm instruments (socket F).	
6.2310.010	Electrolyte KNO ₃ -1 mol/L 250 mL Electrolyte solution KNO ₃ 1 M (reference electrolyte for combined silver electrodes)	
6.0726.100	Ag/AgCl reference electrode (length 12.5 cm) Silver/silver chloride reference electrode with Double-Junction System, installation length 10 cm. The standard ground-joint 14/15 enables easy assembly, and the flexible ground-joint diaphragm, which is insensitive to contamination, can be replaced at any time. The reference electrolyte and the bridge electrolyte can be selected freely according to use, and are easy to replace. This sensor is supplied without electrolyte filling.	

6.0750.100

LL ISE reference electrode

Silver / silver chloride reference electrode with double junction system.

This reference electrode is well suited for:

- automated applications
- ion measurements
- surfactant titrations

The ground-joint diaphragm, which is insensitive to contamination, offers a constant and reproducible electrolyte outflow. Additionally, the reference electrolyte is gelified for even better signal stability. The sensor is delivered with $c(\text{KCl}) = 3 \text{ mol/L}$ as bridge electrolyte, which can be freely selected and exchanged as needed.



6.02104.000

Electrode cable plug-in head G / plug P, 0.55 m

Cable for connecting an analog electrode (plug-in head G) to an analog measuring module from OMNIS.

