



Conductivity measuring cell $c = 0.1 \text{ cm}^{-1}$ with Pt1000 (fixed cable, 0.65 m)

6.00921.040

Conductivity measuring cell made of stainless steel with cell constant $c = 0.1 \text{ cm}^{-1}$ (guide value), with integrated Pt1000 temperature sensor and with fixed cable (0.65 m) for connecting to the OMNIS Measuring Module Conductivity.

This sensor is suitable for measurements of low conductivities (0 to 300 $\mu\text{S}/\text{cm}$), e.g. in deion. water or for measurements in accordance with USP 645>.

Below, the accessories are grouped into Scope of delivery and Optional accessories.

Please keep this printout at hand for ordering replacement material.

These lists may be subject to change.

Scope of delivery 6.00921.040

Qt.	Order no.	Description
-----	-----------	-------------

20 Stk

6.1236.050

Sleeve with SGJ 14/12 mm

Sleeve with SGJ 14/12 mm, polyethylene.



Optional accessories

Order no.	Description	
6.2324.010	Conductivity standard 100 µS/cm 250 mL Conductivity standard for calibration of conductivity measuring cells with cell constant = 0.1/cm.	
6.2324.110	Conductivity standard 100 µS/cm, 5 x 30 mL Conductivity standard for calibration of conductivity measuring cells with cell constant = 0.1/cm.	
6.02101.020	Measuring module conductivity Measurement channel for one OMNIS Titrator or Titration Module for the connection of conductivity measuring cells.	