



5-ring conductivity measuring cell $c = 1.0 \text{ cm}^{-1}$ with Pt1000 (fixed cable 2 m)

6.0920.130

5-ring conductivity measuring cell with cell constant $c = 1.0 \text{ cm}^{-1}$ (guide value), with integrated Pt1000 temperature sensor and fixed cable (2.0 m) for connecting to an 856 Conductivity Module in combination with a sample changer.

This sensor is suitable for automated measurements of medium conductivities ($5 \mu\text{S/cm}$ to 100 mS/cm), e.g., in:

- drinking water
- surface water
- wastewater

Below, the accessories are grouped into Scope of delivery and Optional accessories.

Please keep this printout at hand for ordering replacement material.

These lists may be subject to change.

Optional accessories

Order no.	Description
-----------	-------------

6.2324.110 Conductivity standard 100 $\mu\text{S}/\text{cm}$, 5 x 30 mL

Conductivity standard for calibration of conductivity measuring cells with cell constant = 0.1/cm.



6.2324.010 Conductivity standard 100 $\mu\text{S}/\text{cm}$ 250 mL

Conductivity standard for calibration of conductivity measuring cells with cell constant = 0.1/cm.



6.2301.060 KCl ion standard 250 mL

KCl ion standard ($c(\text{ion}) = 0.1000 \pm 0.0005 \text{ mol/L}$), can also be used as standard solution for conductometry 12.87 mS/cm (25 °C).

