



Conductivity measuring cell $c = 0.1 \text{ cm}^{-1}$ with Pt1000 (fixed cable, 2.0 m)

6.00922.040

Conductivity measuring cell made of stainless steel with cell constant $c = 0.1 \text{ cm}^{-1}$ (guide value), with integrated Pt1000 temperature sensor and with fixed cable (2.0 m) for connecting to the OMNIS Measuring Module Conductivity.


This sensor is suitable for automated measurements of low conductivities (0 to 300 $\mu\text{S}/\text{cm}$), e.g. in deion. water or for measurements in accordance with USP 645>.

Phạm vi giao hàng 6.00922.040

Qt.	Order no.	Description
10 Stk	6.1236.050	Sleeve with SGJ 14/12 mm
		Sleeve with SGJ 14/12 mm, polyethylene.



Phụ kiện tùy chọn

Order no.	Description	
6.2324.010	Conductivity standard 100 $\mu\text{S}/\text{cm}$ 250 mL Conductivity standard for calibration of conductivity measuring cells with cell constant = 0.1/cm.	
6.2324.110	Conductivity standard 100 $\mu\text{S}/\text{cm}$, 5 x 30 mL Conductivity standard for calibration of conductivity measuring cells with cell constant = 0.1/cm.	
6.02101.020	Measuring module conductivity Measurement channel for one OMNIS Titrator or Titration Module for the connection of conductivity measuring cells.	