



5-ring conductivity measuring cell $c = 0.7 \text{ cm}^{-1}$ with Pt1000 (fixed cable 0.65 m)

6.00925.100

5-ring conductivity measuring cell with cell constant $c = 0.7 \text{ cm}^{-1}$ (guide value), with integrated Pt1000 temperature sensor and with fixed cable (0.65 m) for connecting to the OMNIS Measuring Module Conductivity.

This sensor is suitable for measurements of medium conductivities ($5 \mu\text{S}/\text{cm}$ to $20 \text{ mS}/\text{cm}$), e.g., in:

- Drinking water
- Surface water
- Waste water

Phạm vi giao hàng 6.00925.100

Qt.	Order no.	Description
-----	-----------	-------------

20 Stk

6.1236.050

Sleeve with SGJ 14/12 mm

Sleeve with SGJ 14/12 mm, polyethylene.



Phụ kiện tùy chọn

Order no.	Description	
6.2324.010	Conductivity standard 100 $\mu\text{S}/\text{cm}$ 250 mL Conductivity standard for calibration of conductivity measuring cells with cell constant = 0.1/cm.	
6.2324.110	Conductivity standard 100 $\mu\text{S}/\text{cm}$, 5 x 30 mL Conductivity standard for calibration of conductivity measuring cells with cell constant = 0.1/cm.	
6.2301.060	KCl ion standard 250 mL KCl ion standard ($c(\text{ion}) = 0.1000 \pm 0.0005 \text{ mol/L}$), can also be used as standard solution for conductometry 12.87 mS/cm (25 °C).	
6.02101.020	Measuring module conductivity Measurement channel for one OMNIS Titrator or Titration Module for the connection of conductivity measuring cells.	