



KFT equipment for OMNIS Titrator with Polytron

6.05609.010

Complete accessory set for volumetric Karl Fischer titration with Polytron.

Phạm vi giao hàng 6.05609.010

Qt.	Order no.	Description
-----	-----------	-------------

1 PCS	6.00346.100	Double Pt ring electrode for volumetry
-------	-------------	--

Indicator electrode for Karl Fischer titration which, thanks to its two fused platinum rings, is more robust and easier to clean than the conventional platinum pin electrode (6.0338.100).



1 PCS

6.01406.010

Adsorber tube with lid

Adsorber tube with lid for OMNIS Karl Fischer titrator, volumetric



1 PCS

6.01406.220

20 – 90 mL titration vessel

20 – 90 mL titration vessel with screw top for use with 6014050X0 titration vessel lids.



1 PCS

6.01406.250

50 – 150 mL titration vessel

50 – 150 mL titration vessel with screw top for use with 6014050X0 titration vessel lids.



1 PCS

6.01408.010

Set of seals for Karl Fischer titration cell

Complete set of seals, suitable for 6.05609.000 and 6.05609.010.



1 PCS

6.01543.120

Antidiffusion tip

Titration tip with antidiffusion valve, suitable for OMNIS Karl Fischer titrator.



1 PCS

6.02104.040

Electrode cable plug-in head G (pol.) / plug P, 0.55 m

Cable for connecting an analog polarizable electrode (plug-in head G) to an analog measuring module from OMNIS.



3 PCS

6.02709.010

Stopper

Plastic stoppers, suitable for 2.1001.4220 OMNIS Karl Fischer Titrator.



1 PCS

6.02709.020

Septum stopper

Plastic septum stopper, suitable for 6014050X0 titration vessel lids.



2 PCS

6.02709.030

Screw nipple

Plastic screw nipple, suitable for 6014050X0 titration vessel lids.



1 PCS

6.02801.000

Cotton for adsorber tube

Wad of cotton (4 pieces), suitable for OMNIS 601406010 adsorber tube.



1 PCS

6.1448.010

Septum 12 mm, 5 pieces

Set of 5 items.



1 PCS

6.2412.000

Weighing boats

Weighing boat made of glass for the 6.1414.030 titration vessel lid or the 6.1455.31X, 6.1464.32X, 6.1465.320 titration vessel. For KF water determinations. Including protective tube.



2 PCS

6.2816.020

Syringe 10 mL with Luer connection

For various applications in IC and VA



2 PCS

6.2816.030

Needle with Luer connector

Needle for filling in mercury into the Multi-Mode Electrode and for sample addition in Karl Fischer titration.



1 PCS

6.01405.040

OMNIS titration vessel lid for 601406XXX titration vessels for Polytron

Titration vessel lid for volumetric Karl Fischer titration with an OMNIS titrator with Polytron.



1 PCS

6.02008.010

Polytron holder

For fastening the Polytron in place, suitable for 601405040.



For fixation of the Polytron in the 601405040 titration vessel upper part.



Phụ kiện tùy chọn

Order no.	Description
2.136.0100	Polytron PT 1300 D Polytron PT 1300 D - Metrohm version Homogenizer that can be controlled directly by OMNIS Software, tiamo™ or Touch Control. The Polytron PT 1300 D is comprised of a control unit and a drive. The coupling system on the drive permits simple and rapid replacement of the aggregates without additional tools. Solid samples can be pulverized without difficulty. However the instrument is also highly suitable for ensuring thorough mixing in viscous samples.
6.9012.000	Dispersing aggregate for Polytron, 125 mm Standard dispersing aggregate for Polytron 1300 D
6.9012.010	Dispersing aggregate for Polytron, 157 mm Special dispersing aggregate (Metrohm version) with protruding knives



6.02709.040

Stopper

Stopper for titration vessel lids 6014050X0.



6.02711.000

Spoon for paste

Suitable for use in Karl Fischer water content determinations of pasty substances such as honey, cosmetics, greases or various dairy products.

The spoon for paste facilitates the addition of viscous samples. It is handled like any other spatula and enables convenient, rapid, and loss-free addition of pastes into the Karl Fischer titration vessel.

We furthermore recommend the use of a homogenizer for dissolving viscous samples quickly and completely.

Compatible with 6014050X0 titration vessel lids.



6.2811.000

Molecular sieve

Molecular sieve. Bottle containing 250 g. Pore size: 0.3 nm. Without moisture indicator. For Rancimats and Karl Fischer instruments.



6.2811.010

Molecular sieve (US)

Molecular sieve. Bottle containing 250 g. Pore size: 0.3 nm. Without moisture indicator. For Rancimats and Karl Fischer instruments.

This product is intended for all users who are US residents. The material is rated hazardous in the US in accordance with OSHA Hazard Communication Standard (29 CFR 1910.1200).

