



5-ring conductivity measuring cell $c = 1.0 \text{ cm}^{-1}$ with Pt1000 (fixed cable 2 m)

6.0920.130

5-ring conductivity measuring cell with cell constant $c = 1.0 \text{ cm}^{-1}$ (guide value), with integrated Pt1000 temperature sensor and fixed cable (2.0 m) for connecting to an 856 Conductivity Module in combination with a sample changer.

This sensor is suitable for automated measurements of medium conductivities ($5 \mu\text{S/cm}$ to 100 mS/cm), e.g., in:

- drinking water
- surface water
- wastewater

Phụ kiện tùy chọn

Order no.	Description
-----------	-------------

6.2324.110

Conductivity standard 100 $\mu\text{S}/\text{cm}$, 5 x 30 mL

Conductivity standard for calibration of conductivity measuring cells with cell constant = 0.1/cm.



6.2324.010

Conductivity standard 100 $\mu\text{S}/\text{cm}$ 250 mL

Conductivity standard for calibration of conductivity measuring cells with cell constant = 0.1/cm.



6.2301.060

KCl ion standard 250 mL

KCl ion standard ($c(\text{ion}) = 0.1000 \pm 0.0005 \text{ mol/L}$), can also be used as standard solution for conductometry 12.87 mS/cm (25 °C).

