



CVS electrode equipment with 3 mm platinum electrode for Professional VA/CVS instruments

6.5339.020

Complete electrode set for the determination of organic additives in electroplating baths with CVS (Cyclic Voltammetric Stripping). Contains rotating platinum working electrode (electrode diameter: 3 mm), reference electrode, platinum auxiliary electrode, measuring vessels and electrolyte solutions.

Phạm vi giao hàng 6.5339.020

Qt.	Order no.	Description
1 PCS	6.0343.100	Separate Pt Rod Electrode
		Platinum auxiliary electrode for voltammetry and CVS.



1 PCS

6.0728.130

LL-Ag/AgCl reference electrode for CVS

Inner reference system with electrolyte $c(\text{KCl}) = 3 \text{ mol/L}$. For use with electrolyte vessel made of plastic 6.1245.010.



1 PCS

6.1204.170

3 mm platinum electrode tip for CVS

Metal electrode tip made of platinum (Pt), diameter of electrode disk 3 mm, polished surface, M3 thread connection. Temperature range: 0 - 40 °C.



1 PCS

6.1204.510

Driving axle for rotating disk electrode (RDE)

Driving axle for rotating disk electrode (RDE) with titanium axle for Professional VA/CSV instruments, M3 thread.



1 PCS

6.1245.010

Electrolyte vessel made of plastic

Electrolyte vessel made of plastic with ceramic diaphragm for double junction reference electrode.



1 PCS

6.1415.210

Titration vessel / 10-90 mL



1 PCS

6.1415.250

Titration vessel / 50-150 mL



1 PCS

6.2308.020

Electrolyte 3 mol/L KCl (250 mL)

Electrolyte solution $c(\text{KCl}) = 3 \text{ mol/L}$, (for Ag/AgCl reference systems)

1 PCS

6.2310.010

Electrolyte KNO_3 -1mol/L 250 mLElectrolyte solution KNO_3 1 M (reference electrolyte for combined silver electrodes)