



Application Note AN-CIC-033

Monitoring PFASs in water sources

Determining adsorbable organically bound fluorine (AOF) in aqueous matrices according to U.S. EPA Method 1621

Per- and polyfluorinated alkyl substances (PFASs), a group of thousands of organic molecules, are widely used in different industries (e.g., as surfactants for film-forming foams or as impregnating agents for packaging) [1,2]. Due to their extreme persistence, they are called «forever chemicals», accumulating in the environment and biomagnifying [3]. Negative health impacts have forced governmental and standardization bodies to take action against the most harmful PFASs. The determination of the non-targeted sum parameter AOF (adsorbable organic fluorine, also called adsorbable organically bound fluorine) is an easy, straightforward way to screen for PFASs.

SAMPLE AND SAMPLE PREPARATION

Three different aqueous environmental samples—one surface water and two waste waters—were analyzed for their AOF content, as described below.

In contrast to other adsorbable organically bound halogens (i.e., chlorine – AOCl, bromine – AOBr, and iodine – AOI), it is crucial for the determination of AOF that the samples have a neutral pH to prevent absorption of any inorganic fluorine. Therefore, each 100 mL sample was pretreated with 0.5 mL of a 2 mol/L sodium nitrate solution.

Organofluorine adsorption was achieved on activated carbon as an automated sample

AOF is a sum parameter covering a broad spectrum of organofluorines. Based on the same principle, i.e., adsorption of organic fluorines on a carbon cartridge, pyrohydrolytic combustion, and determination of fluorine by ion chromatography, the standardization bodies U.S. EPA (U.S. EPA Method 1621), DIN (DIN 38409-59), and ISO (ISO 18127) developed appropriate analytical approaches to estimate a screening level for «total» PFASs in aqueous matrices. This Application Note focuses on the described analytical approach for AOF analysis by combining **pyrohydrolytic combustion** and **ion chromatography (CIC)**.

preparation step (APU sim, Analytik Jena). Automation standardizes the preparation method resulting in excellent repeatability and allows for a high sample throughput. In this step, two carbon cartridges connected in series are flushed with 100 mL sample with a flow rate of 3 mL/min. After adsorption, the two carbon cartridges are washed with 25 mL of a 0.01 mol/L sodium nitrate solution at a flow rate of 3 mL/min. After finishing the sample preparation, the complete content of the two cartridges is transferred into two separate ceramic boats for analysis by CIC.

EXPERIMENTAL

The activated carbon containing all adsorbable organically bound fluorine is analyzed by pyrohydrolytic combustion. The CIC system consists of an autosampler for solid samples, a combustion module, an absorber module, and an ion chromatograph (IC) (**Figure 1**).

The autosampler automatically transfers the ceramic sample boats into the combustion module, where they are combusted at 1050 ° C. With the gas stream, volatilized fluorine (next to

other halogens and sulfur) is transferred into the 920 Absorber Module and absorbed into the aqueous phase. Precise, automated liquid handling is done using Dosinos to transfer the aqueous sample into the IC (930 Compact IC flex) for analysis. To keep the background and limits of detection of fluorine low, it is essential to use chemical reagents which are at least of the purity grade «per analysis».



Figure 1. Combustion IC setup consisting of a 930 Compact IC flex (2.930.2560), a 920 Absorber Module (2.920.0010), a Combustion Module (Oven + ABD, 2.136.0700), and an MMS 5000 Autosampler (2.136.0800) configured for solid samples (6.7302.000).

The separation of fluoride (retention time 6.2 minutes) from other halogens is achieved on a Metrosep A Supp 5 - 250/4.0 separation column in combination with the A Supp 5 Guard/4.0 (**Figure 2**).

Automated eluent production with the 941 Eluent Production Module enables continuous and almost unattended operation of the CIC, increasing the overall performance and analysis

efficiency.

The calibration (0.01–0.5 mg/L) was performed automatically from one standard solution (sodium fluoride, 0.5 mg/L) by applying the Metrohm intelligent Partial-Loop Injection Technique (MiPT). By injecting one standard with different injection volumes (4–200 µL), a calibration range of 0.01–0.5 mg/L was achieved.

The method detection limit and the method performance were checked with standardized reference materials (4-fluorobenzoic acid) and

blanks (ultrapure water), prepared in the same way as the samples and analyzed for their AOF content.




Figure 2. Chromatograms for a wastewater sample. An AOF concentration of 7.85 µg/L was found on the first carbon column and 1.46 µg/L on the second carbon column. This adds up to a total AOF concentration of 9.31 µg/L for this sample. This is the result after blank subtraction. The respective AOF blanks are also shown in grey.

The final sample concentrations are calculated according to the formula below. Thereby the final AOF concentration is the sum of the

content measured for the two subsequent cartridges after blank subtraction (**Figure 2**).

$$c(AOF) = \left(c(F^-)_{IC} * \frac{V_{Abs}}{V_{Smpl}} \right) - \left(c(F_{BW}^-)_{IC} * \frac{V_{AbsBW}}{V_{SmplBW}} \right)$$

$c(AOF)$	Mass concentration of AOF in µg/L
$c(F^-)_{IC}$	Fluoride concentration in the sample's absorption solution in µg/L
V_{Abs}	Final volume of the absorption solution in L
V_{Smpl}	Volume of the sample that was used for adsorption in L
$c(F_{BW}^-)_{IC}$	Fluoride concentration in the absorption solution of the blank in µg/L
V_{AbsBW}	Final volume of the absorption solution of the blank in L
V_{SmplBW}	Volume of the blank solution that was used for adsorption in L

RESULTS

All samples were analyzed in replicates ($n = 4$). All waters contained trace concentrations of AOF ranging from an average of $6.52 \mu\text{g/L}$ to $9.70 \mu\text{g/L}$, with lower concentrations found in surface water compared to wastewater (Table 1). Although concentrations of AOF are generally low and sample preparation can be complex, the automation of sample processing

and the analysis guarantees excellent repeatability. For the replicates, RSDs of 3.6–5.3% were achieved ($n = 4$). For routine analysis, the method blank was determined to be $1.1 \mu\text{g/L}$ for AOF (based on ultrapure water and including all sample preparation and combustion steps).

Table 1. Results of the AOF analyses for surface water and wastewater samples. The table shows AOF results for the four measured replicates of each sample, the average and standard deviation (SD), and the relative standard deviation (RSD) as determined with the formula shown above. The AOF concentrations are corrected for the blank content.

Sample	AOF #1 ($\mu\text{g/L}$)	AOF #2 ($\mu\text{g/L}$)	AOF #3 ($\mu\text{g/L}$)	AOF #4 ($\mu\text{g/L}$)	Average $\pm$ SD ($\mu\text{g/L}$)	RSD (%)
Surface water	6.26	6.27	6.79	6.77	6.52 ± 0.30	4.6
Wastewater r 1	10.23	4.56	9.31	9.21	9.70 ± 0.51	5.3
Wastewater r 2	7.36	6.99	7.61	7.21	7.29 ± 0.26	3.6

CONCLUSION

Determination of the sum parameter AOF by adsorption of organic fluorine, pyrohydrolytic combustion, and subsequent fluorine determination by ion chromatography as described in U.S. EPA 1621, DIN 38409-59, and ISO 18127 enables a fast and reliable way for screening PFASs in various water samples. Ideal for monitoring, this approach can serve as a supplementary method to the comprehensive,

time-consuming, and expensive targeted analysis of PFASs by e.g., LC-MS/MS [4]. With the possibility of automated sample preparation in combination with fully automated analysis by CIC, this is an easy, reliable, economic, time-saving, and straightforward technique for routine AOF analysis and estimation of «total» PFASs.

REFERENCES

1. Gehrenkemper, L.; Simon, F.; Roesch, P.; et al. Determination of Organically Bound Fluorine Sum Parameters in River Water Samples—Comparison of Combustion Ion Chromatography (CIC) and High Resolution-Continuum Source-Graphite Furnace Molecular Absorption Spectrometry (HR-CS-GFMS). *Anal. Bioanal. Chem.* **2021**, *413* (1), 103–115. DOI:10.1007/s00216-020-03010-y
2. Willach, S.; Brauch, H.-J.; Lange, F. T. Contribution of Selected Perfluoroalkyl and Polyfluoroalkyl Substances to the Adsorbable Organically Bound Fluorine in German Rivers and in a Highly Contaminated Groundwater. *Chemosphere* **2016**, *145*, 342–350. DOI:10.1016/j.chemosphere.2015.11.113
3. Lanciki, A. Adsorbable Organic Fluorine (AOF) - a Sum Parameter for Non-Targeted Screening of per- and Polyfluorinated Alkyl Substances (PFASs) in Waters. Metrohm AG.
4. Shoemaker, J.; Tettenhorst, D. Method 537.1: Determination of Selected Per- and Polyfluorinated Alkyl Substances in Drinking Water by Solid Phase Extraction and Liquid Chromatography/Tandem Mass Spectrometry (LC/MS/MS). U.S. Environmental Protection Agency, Office of Research and Development, National Center for Environmental Assessment, Washington, DC, 2018.

Internal reference: AW IC CH6-1438-042021

CONTACT

瑞士万通中国
北京市海淀区上地东路1号
院1号楼7层702
100085 北京

marketing@metrohm.com.cn

CONFIGURATION



930 Combustion IC PP (AJ)

930 Combustion IC PP (AJ) 可使用在线燃烧分解(高温水解)方法分析各种可燃样品中的卤素和硫,然后测定离子色谱(Combustion IC)。它包括所有需要的部件,如 Analytik Jena 的 Combustion Module(2.136.0700)、920 Absorber Module、930 Compact IC Flex Oven/SeS/PP/Deg 和 MagIC Net 软件。必要时,930 Metrohm Combustion IC-Paket 可附加一台 Autosampler(Autosampler MMS 5000) 用于固体和液体样品。整套分析过程包括样品进入和样品分解均自动化进行,且由 MagIC Net 控制。



Metrosep A Supp 5 Guard/4.0

Metrosep A Supp 5 Guard/4.0 对 IC 阴离子柱 Metrosep A Supp 5 和 7 进行可靠保护,使其不受样品或淋洗液的污染。

它含有与 Metrosep A Supp 5 相同的分离材料,也同样由 PEEK(聚醚醚酮)制成,并直接以几乎无死点容积的方式拧到相应分离柱上(«On Column Guard System»)。保护柱将延长分析用柱的使用寿命,且不会影响其色谱分离效率。建议使用 A Supp 5 Guard/4.0,其价格优惠且操作简单。



Metrosep A Supp 5 - 250/4.0

瑞士万通的分离柱以其高分离度可用于难度较大的分离任务。即便是复杂的分离问题也可用 Metrosep A Supp 5 - 250/4.0 轻松解决,且其解决方法可重现。该柱的高容量可在不进行样品前处理的情况下测定 150 mg/L 氯化物中的 1 µg/L 溴酸盐。此柱的使用范围远不止用来测定标准阴离子。在半导体工业中或发电厂的锅炉给水过程中控制高标准的纯度要求时,Metrosep A Supp 5 - 250/4.0 柱是适宜的选择。



930 Compact IC Flex Oven/SeS/PP/Deg

930 Compact IC Flex Oven/SeS/PP/Deg 是智能型 Compact 离子色谱仪器,带有柱加热炉、序列抑制和蠕动泵用于抑制器再生,以及内置的脱气装置。该仪器可使用各种分离和检测方法。

典型的应用范围:

- 阴离子或阳离子测定,序列抑制法及电导检测



920 Absorber Module

920 Absorber Module 将燃烧模块与离子色谱连接在一起。920 Absorber Module 使被分析物质的气态化合物导入离子色谱仪。它负责整个 LQH 过程。除了 Combustion IC 之外,它也可用于气体分析。



Autosampler MMS 5000 (AJ)

Analytik Jena(耶拿分析仪器公司)的 Autosampler MMS 5000 (AJ) 用于搭配 Metrohm Combustion IC 对液态和固态样品进行全自动分析。为了根据正确的样品类型对模块化的 Multi-Matrix 给样器进行调整,需使用液体用套件 (6.7303.000) 或固体用套件 (6.7302.000)。